

## Evaluation of an In-House PCR Assay for the Detection of *Neisseria Gonorrhoeae*

Zodidi Dilinga\*, Nathlee S Abbai and Khine Swe-Swe Han

School of Clinical Medicine Research Laboratory, Nelson R Mandela School of Medicine, University of KwaZulu-Natal, Durban, South Africa.

Published November 28, 2023

### ABSTRACT

**Introduction:** Sexually transmitted infections (STIs) are a major public health concern worldwide, affecting quality of life and causing serious morbidity and mortality. *Neisseria gonorrhoeae* is challenging to diagnose by culture due to its fastidious nature. A rapid, highly sensitive PCR assay is urgently needed for routine *N. gonorrhoeae* detection to prevent widespread infection and reproductive complications. The aim of this study was to evaluate the sensitivity of an in-house PCR assay specifically for the detection of *N. gonorrhoeae*.

**Methods:** For the design of the in-house PCR, the *16S rRNA* of a *N. gonorrhoeae* isolate was accessed from GenBank (AJ247239.2). Using a primer design tool (<https://www.ncbi.nlm.nih.gov/tools/primer-blast/>), the primers were selected and used in a qPCR SYBR green assay. The in-house *16S rRNA* assay was compared to the commercially available TaqMan assay which uses primers and probes specific for *N. gonorrhoeae*. The *16S rRNA* primers were tested for cross reactivity towards six *Neisseria* strains. The DNA from the following *Neisseria* isolates: *N. elongate*, *N. cinerea*, *N. weaver* and *N. sicca* were tested.

**Results/Findings:** The analysis showed successful *16S rRNA* amplification in all clinical *N. gonorrhoeae* isolates. Positive amplification matched in-house and TaqMan results, displaying strong correlation. Cross-reactivity analysis confirmed assay specificity for *N. gonorrhoeae* detection, excluding non-gonococcal *Neisseria* strains.

**Conclusion:** This study showed that the in-house PCR had successfully detected *N. gonorrhoeae*. The in-house assay showed a good correlation with a commercially available assay. This assay needs to be validated using different clinical samples to show its promise as a diagnostic test.

**Keywords:** *Neisseria gonorrhoeae*, Cross-reactivity, TaqMan assay, In-house PCR assay

**Corresponding author:** Zodidi Dilinga, School of Clinical Medicine Research Laboratory, Nelson R Mandela School of Medicine, University of KwaZulu-Natal, Durban, South Africa, E-mail: zodididilinga@yahoo.com

**Citation:** Dilinga Z, Abbai NS & Swe-Swe Han K. (2023) Evaluation of an In-House PCR Assay for the Detection of *Neisseria Gonorrhoeae*. J Infect Dis Res, 6(S4): 28.

**Copyright:** ©2023 Dilinga Z, Abbai NS & Swe-Swe Han K. This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.